

A contemporary Design Approach for Prostheses

Giovanni Maver and Angelo Magni design prostheses with the help of the CAD/CAMCNC-technique. This technique provides a virtual look at the muscles and bones ratio and digital imagery offers precise design options. This method cuts material costs and saves time.

Nowadays, accurately designing and creating dental prostheses is much easier thanks to modern technology ranging from X-ray templates, drilling templates and immediate prostheses to the final restoration consisting of a chrome-cobalt structure with veneers. Aesthetic and functional harmony are the main focus during the design phase.

Software assistance

The patient is wearing a partial prosthesis with cuspid brackets (Fig. 1). We agree on a software-assisted restoration option to improve aesthetics and functional properties. In order to properly identify the bone and soft tissue ratio (Figs. 2a and 2b), we start by digitalizing the models (Figs. 3a and 3b) and create an X-ray template with a modified occlusal plane (position




Fig. 1 — Initial situation.




Giovanni Maver
Pedrengo, Italy



Angelo Magni
Fornovo, Italy



Figs. 2a and 2b — Bone/muscle situation.



Figs. 3a to 3d — Digitalization of the models and X-ray template.

and size of the teeth) (Figs. 3c and 3d). The template has a radiopaque base (red arrows, Fig. 3d).

TAC Cone Beam

The TAC Cone Beam (Figs. 4a and 4b) visualizes soft tissue and cranial bones (Fig. 5). For an accurate positioning of the cranium, we use virtual reference points (Fig. 6):

- The vertical plane: Tangent from the highest point of

the crista galli (protuberance of the ethmoid bone) to the frontal nose spine;

- The sagittal plane: Tangent from the highest point of the crista galli to the centre of the sella turcica (part of the sphenoid bone); and
- The horizontal plane: Tangent (on both sides) from the lower edge of the eye sockets to the porion (highest point of the bony ear canal opening), also known as the “Frankfort Horizontal.”

By simply importing the working models into the CAD



Figs. 4a to 4b — Patient with an X-ray template. Measuring with the TAC Cone Beam.




Fig. 5 — Image of the hard and soft tissues.

program, we obtain bone structures with the previously digitized models (Fig. 7).

Drilling template

Visualizing the patient's real muscle-bone ratio allows us to evaluate various reference points necessary for determining the occlusal plane. For design of the temporary restoration, we use the Camper's line (the Camper's line lies between the tangents running from the nostril to the lower edge of the ear canal on each side of the face), with the bipupillary line serving as a reference point for the sagittal inclination angle of

the frontal plane. This line runs parallel to the six front teeth and through the centres of the pupils (Fig. 8).

We now proceed by determining the size and positioning of the teeth with the help of biovolumetric templates (used for designing permanent restorations) (Figs. 9a and 9b). Determining the accurate bone amount and quality, the precise implant positions and proper positioning of the teeth is crucial in this planning phase (Fig. 10). The next step is the fabrication of drilling templates, temporaries and models with implant positions (Figs. 11a to 11f). The therapeutic temporary prostheses with gingiva and guides are ready to be placed on the models once the abutments have been properly integrated (Fig. 12).




Fig. 6 — Positioning of the cranium.




Fig. 7 — TAC and inserted models.



Fig. 8 — Camper's plane and bipupillary line for the occlusal plane.



Calls2Wealth is a system developed for dental professionals like you, to grow your practice by attracting more clients, increasing your case acceptance, elevating your clients' satisfaction and of course increasing your bottom line.

TAKE ACTION NOW! - call C2W at 905-929-8279

Denturists - Are *YOU* getting the **MO\$T** out of your **IMPLANT SUPPORTED CASES?**
Get more control in **4 EASY STEPS!**

1.



Your Patient - Your Case
It begins & ends with **YOU!**

2.



As part of the Team,
Dr. Estrabillo provides your patient
with **Implant Surgical Services**

3.





Our Technical Team provides you with
custom designed components
so that you can finish your own cases!

4.



YOU complete the case!



Figs. 9a and 9b — Biovolumetric template for planning the teeth's correct position and size.





Fig. 10 — Optimization with consideration of the present bone structure, implant and positioning of the teeth.



Implantation and Temporaries



Prior to inserting the implants and applying the therapeutic temporaries, we start with the positioning of the drilling templates (Figs. 13a to 14c).

This is a minimally invasive procedure with little or no bone structure damage and the patient receives a complete restoration the same day. After the implant sites have healed

properly, the next step is the planning and construction phase for the permanent restoration.

Permanent prostheses

The CAD technique allows modification of previously designed work at any time. Individual harmony derives from aesthetics in combination with functional balance, such as



Figs. 11a to 11f — Drilling templates, temporaries and models with implants.





Fig. 12 — Temporaries with enamel, gingiva and guides.


chewing, swallowing and speaking. Adequate space is, therefore, an important factor (Fig. 15). Since different dental teams might work on the prosthesis' design and construction, for organizational purposes we arranged a standard series of instruments for the project (?). The biovolumetric templates were created because of the need to record codified working methods based on consistent measures and proportions.

Spacing conditions

In a physiological oral situation, certain areas of the teeth, marked with red dots in Fig. 16, display a connection pattern. Figures 17 and 18 show different shapes of tooth arches and teeth. The red dots connected by lines denote the teeth that will be preserved. We are also able to identify the individual



Figs. 13a to 13d — Drilling template and fitting of the temporaries.



Figs 14a to 14c — Before and after the procedure.

Fig. 15 — Space condition.




Fig. 16 — Position of the teeth.

cusps of the posterior teeth and the central growth cones of the anterior teeth (Fig. 19). Space varies, depending on the biotype and size of the individual tooth. The biovolumetric templates (Fig. 20) serve as organizational and structural tools. They enable us to systematically place the prostheses, thereby preventing unnecessary “empty” spaces.

Individualization

A facial analysis is useful for creating aesthetical and functional harmony. Face classification (Fig. 21a) determines biotype,

asymmetry, typology and oral tissue structures. I then select the proper biovolumetric template (Fig. 21b) and the matching tooth shape and continue by determining the Spee and Wilson curve, inclination angle, dominance, position and the anterior teeth set-up with the help of facial asymmetry and oral tissue structures.

Data maintenance

After the healing period, the next step is to digitalize the models of the temporary prosthesis and the new models with all available data (models, TAC etc). The




Fig. 17 — Dental arch shapes and matching facial shapes.




Fig. 18 — Arch/teeth dimension..




Fig. 19 — Cusps/position ratio




Fig. 20 — Biovolumetric templates, upper plane and bottom calotte.



Fig. 21a — Face classification.




Fig. 21b — Facial shape for selecting the correct biovolumetric template and teeth.



Fig. 21c — Asymmetry of the face, tissue structures in the oral area for the set-up of the anterior teeth.

following steps provide valuable information to identify or create available space:

- Take the size of the biovolumetric template (Fig. 22) using the proportional code of the Wheeler chart in accordance with the preserved natural teeth (Fig. 23), thus generating an accurate template for the patient.
- Search the program for the correct teeth which fit the template in terms of size and shape (Fig. 24) and create permanent teeth for the patient.
- Position the templates for the occlusal plane in relation to maxilla and mandible using the available bone and muscle information (Fig. 25). The Camper's line (Fig. 26) and the Frankfort Horizontal line (Fig. 27) serve as reference points.

There is no occlusal plane per se, but rather an imaginary surface where the upper and lower teeth make contact. The occlusal plane is not a tangible element; it serves merely as a reference point during the design process. Determining a plane requires three reference points, but which points do we use?

During second dentition, the lower first molars and the lower incisors break through almost simultaneously and between these teeth lies the "occlusal plane." The two




Fig. 22 — Biovolumetric templates.



Fig. 23 — Fitting the biovolumetric template with the natural teeth.



Fig. 24 — Adjusting shape and size of the teeth.




Fig. 25 — Positioning of the templates in relation to the bone and muscle situation in the maxilla and mandible.




Fig. 26 — Camper's plane, reference point for the top template.




Fig. 27 — Frankfort Horizontal line, reference point for the bottom template.




Fig. 28 — Reference planes.




Fig. 29 — Spee's curve, Wilson's curve and Crista Galli.

mesial-lingual cusps of the first upper molars serve as a reference point in the posterior area and the mesial-incisal point between the lower number ones (incisor point) are used for reference in the anterior area (Fig. 28). We now proceed by adding the following information:

- Integration of the Spee's and Wilson curve (Fig. 29), according to the face asymmetry;
- Dominance/set-up of the anterior teeth in relation to the oral area; and
- Completing the design process in the CAD software (Figs. 30 and 31).

Advantages

Work procedures, reference points and proportions are integrated into the software easily and are accessible for any future rehabilitation work with similar characteristics. There is a difference between an empty piece of paper (Fig. 32) and a coded number puzzle (Fig. 33). By simply connecting the dots, we are able to create a basic drawing, but creating work with aesthetical and functional harmony for the patient depends mostly on the technician's skills.

Finishing

A prototype made of acrylic allows us to evaluate the aesthetics (Fig. 35a). After a try-in and necessary modifications, we move on to the planning and finishing of the framework (made of a chrome-cobalt alloy) for the permanent prosthesis (Fig. 35b). We fabricate another acrylic temporary prosthesis (Figs. 35c and 35d) prior to applying veneering to the framework. After firing, there is another try-in. Once the Camper's plane and aesthetics in the anterior region are verified (Fig. 36), the prostheses are ready for finishing (Figs. 37a to 37c).




Fig. 30




Fig. 31


Figs. 30 and 31 — The finished design.



Fig. 32 — Adequate space for the design.



Fig. 33 — Reference points, proportions.



Work assessment

Achieving aesthetic facial harmony depends on the interaction of various aspects which define facial expressions. Functional harmony is created by the proper interaction of prosthesis and muscles. The quality of the occlusal contacts which stabilize centric and impact dynamics is remarkable. The chewing cycle is the physiological component of functional harmony, demonstrated by the





Fig. 34 — Base design.



Figs. 35a to 35d — The prototype, reduced framework, in acrylic for the aesthetic evaluation.



Fig. 36 — Camper's plane and Frankfort Horizontal line/anterior aesthetics.




opening and closing movements during the masticatory process.

For the verification and evaluation of masticatory movements, chewing waxes made by Kerr and a banana are ideal. By simply placing chewing wax on the occlusal surfaces and letting the patient chew a piece of banana, we are able to evaluate the static and dynamics of the occlusal contacts at the same time (Fig. 39). The soft



Figs. 37a to 37c — Finished prostheses.



Figs. 38a to 38i — The smiling patient from different views.

texture of the banana does not alternate the wax (Fig. 40). After verifying the masticatory process, we then test the habitual closing movement (Fig. 41). Verification of the quality of the occlusal contacts, according to the occlusal concept, is performed with occlusal paper.



Fig. 39 — Trial chewing with a banana.



Fig. 40 — Chewing and swallowing.



Fig. 41 — The myocentric.

Summary

Immediate prostheses:

- Digitalization of the initial situation;
- Planning and fabrication of the X-ray template on the TAC Cone Beam;
- Designing the drilling templates; and
- Positioning of the implants and insertion of the immediate prostheses.

Permanent prostheses (without TAC):

- Digitalization of the models after healing period and maxillomandibular relationship record;
- Collecting data for ensuring individual aesthetic and functional harmony;

- Creating biovolumetric templates;
- Individualizing aesthetic and functional harmony;
- Fabricating the prototype for aesthetic evaluation;
- Milling the framework and preparing it for veneering; and
- Applying aesthetic veneers to the framework.

Conclusion

Modern technology provides the necessary tools to preserve a patient's muscular and bone structures and, maybe eventually, their functional masticatory movements. Working with biovolumetric templates facilitates and speeds up the planning and production process, but the quality of the work depends on the technician's skills and experience. ■



BIODEGRADABLE

DentaNet
MOUSSE

**Gentle non-abrasive daily
mousse cleanser**

For all types of dentures

PREVENTS:

- ✓ stains
- ✓ bacteria
- ✓ limestone
- ✓ tartar

NEW

**ALSO
AVAILABLE:**

Liquid Dentanet
Super concentrated
for soaking







73 St-Louis, Lemoyne (Québec) J4R 2L3
T 450-923-3368 | 1-877-781-8854 | F 450-923-8466
dentanet2000@hotmail.com | www.dentanet2000.com